

HYGITECH

Dry Block incubator

en - INSTRUCTIONS FOR USE:

Dry Block incubators are calibrated to incubate biological indicator products at pre-set temperature ranges. The incubator's intended design is to be plugged in and left on. This insures that the temperature remains stable and eliminates variables with the biological indicator results.

Installation

1. Remove incubator from packaging and inspect for damage.
2. Plug incubator into an earthed or grounded outlet, away from any water sources.
3. Follow the manufacturer instructions with regards to incubation times for the biological indicator. Prior to incubation, be sure to activate the biological indicator (where applicable) as directed by the manufacturer.
4. Incubator is intended to be plugged in and left on.

Operational Precautions

- Incubator is designed for indoor use only and should be placed where the grounded power source is easily assessable. Do not place incubator in drafty areas, near open windows, air conditioning unit, heating vents, or heat-generating equipment since fluctuating temperatures may affect the incubator temperature.
- Keep away from water sources and never remove bottom cover.
- To clean the unit, unplug incubator, dampen cloth with water and wipe to remove residuals.

Note: If the equipment is used in a manner not specified by the manufacturer, the protection provided by the equipment may be impaired. Never use vial crusher with a glass tube or ampule.

Specifications

- Catalog Number: NDB-060-E
- Size: 3.75" x 5.25" x 3"
- Weight: 2 lbs
- Number of wells: 15
- Electrical Rating: 240VAC, 50 Hz frequency, 0.375A
- Ambient operation room temp. 5o C to 40o C. Maximum relative humidity 80%
- Altitude < 2000 M.
- Pollution Category 2
- Installation Category 2
- Main power supply voltage fluctuation not to exceed +/- 10% of the nominal voltage
- Safety: Manufactured in accordance to CSA C22.2 No 61010-1/UL61010-1 & IEC/EN61010-1 Electrical Safety Standards.
- Temperature shall remain constant within its stated temperature range.

Version 1.0 – 04/2018