

TECHNICAL SHEET

Designation	Desinfectant Wipes Medibase				
Reference(s)	The box of 100 v		3131820		
	The refill bag of 10	0 wipes	3131821		
Product picture			Medibase Miscopaled Wipes Mi		
Product description	Wines impregnated with	a disinfectant solution for c	eaning and disinfecting non-		
Product description	•				
	invasive and non-submersible medical devices (small medical and dental equipment).				
Characteristics and	Less than 5% non-ionic surfactants, perfumes including limonene, disinfectants:				
compositon	Ethanol, Alkyl, Dimethylbenzylammonium chloride, Propan-2-ol.				
•					
	Wipes impregnated with a disinfectant solution for cleaning and disinfecting non-				
	invasive and non-submersible medical devices (small medical and dental				
	equipment).				
	 Wipes particularly large, thick, strong and sweet. Less than 5% non-ionic surfactants, perfumes including limonene, disinfectants Ethanol, Alkyl, Dimethylbenzylammonium chloride, Propan-2-ol. 				
	ACTIVITE N	MICROBIENNE	Contact time		
	General treatments				
	Bactericidal	EN13727* - EN14561	5min		
		EN14561: MRSA	5min		
		EN14562: Aspergillus	5min		
	Fungicidal	fumigatus EN13624 - EN 14562:			
	Yeasticidal	Candida albicans	5min		
		EN14348 - EN 14563 :	15min		
	Tuberculocidal	Mycobacterium terrae			
	Virucidal	EN14476*: HBV - HCV - Adenovirus Coronavirus - Novovirus - VRS - H1N1 - HSV1			
		EN14476: Polyomavirus	1min		

		T			
		EN14476: Rotavirus	15min		
	Additional use				
	Bactericidal	EN1276	30s		
		EN13697	5min		
	Fungicidal	EN13697: Aspergillus	5min		
	Fungicidal	niger EN1650* - EN 13697:			
	Yeasticidal	Candida albicans	5min		
	*In dirty conditions	Carraraa arbrearre			
Packaging	- The box of 100 wip	es			
	The refill bag of 100	wipes			
Shelf life	3 month from the day of opening				
Indications of use					
	LOT				
	L _				
	Ready-to-use wipes. Single-use wipes. When using the box for the first time:				
	1. Remove the cover,				
	2. Permanently remove the seal,				
	3. Thread the central wipe through the hole in the cover with clean hands				
	and				
	4. Replace the	cover.			
	5. Use the wipe on the equipment.6. Leave on for 30 seconds to 15 minutes according to the antibacterial activity required.				
	7. Leave to dry in the open air.				
	8. Do not rinse	·			
	9. Close the fla	ap after each use to prevent the wip	es from drying out.		
			, -		
Legal information	Medical Device of cl	Medical Device of class IIA under European Directive 93/42/EEC including			
	Directive 2007/47/CE CE Marking 0459. G-MED Notified body.				
	Manufacturer : FELT S.A				

Version 2.0

Updated 2020/10/23 by Marion CHABANAT